

IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

In re the application of:

Jussi NURMI et al.

Serial Number: 10/579,137

Group Art Unit: 1637

Filed: May 15, 2006

Examiner: Mummert, Stephanie K.

For: NUCLEIC ACID AMPLIFICATION ASSAY AND ARRANGEMENT THEREFOR

REQUEST FOR RECONSIDERATION

Commissioner for Patents

P.O. Box 1450

Alexandria, VA 22314

February 11, 2011

Sir:

In response to the Official Action mailed October 12, 2010, a Petition for a one month Extension of Time being submitted herewith, please reconsider this application in view of the following remarks. Claims 18, 19 and 21-30 are pending.

The 35 U.S.C. § 102(e) rejection of claims 18, 19 and 21-30 over U.S. published application No. 2004/0229349 to Daridon is traversed. The claimed assay includes retaining biological particles containing an analyte(s) of interest on a filter, followed by removing the biological particles from the filter by a flush flow in a second direction opposite said first direction, and then analyzing biological particles contained in the flush flow by means of a nucleic acid amplification assay. The flush flow is analyzed for the analyte(s) without any purification.